



Physical-Technical Testing Institute
Ostrava - Radvanice



(1) **Supplementary EU - Type Examination Certificate No.23**

(2) **Equipment or Protective Systems Intended for Use
in Potentially Explosive Atmospheres
(Directive 2014/34/EU)**

(3) EU - Type Examination Certificate number:

FTZÚ 08 ATEX 0340X

(4) Product: **ALTAIR 5X Multigas Detector**

(5) Manufacturer: **MSA – The Safety Company**

(6) Address: **1000 Cranberry Woods Drive, Cranberry Twp, PA 16066, USA**

(7) This supplementary certificate extends EC - Type Examination Certificate No. FTZÚ 08 ATEX 0340X to apply to products designed and constructed in accordance with the specification set out in the Schedule of the said certificate but having any variations specified in the Schedule attached to this certificate and the documents therein referred to.

(8) The Physical-Technical Testing Institute, Notified Body number 1026, in accordance with Articles 17 of Directive 2014/34/EU of the European Parliament and of the Council, dated 26.02.2014, certifies that this product, as modified by this supplementary certificate, has been found to comply with the Essential Health and Safety Requirements relating to the design and construction of products intended for use in potentially explosive atmospheres given in Annex II to the Directive.

(9) In accordance with Article 41 of Directive 2014/34/EU, EC-Type Examination Certificates referring to 94/9/EC that were in existence prior to the date of application of 2014/34/EU (20.04.2016) may be referenced as if they were issued in accordance with Directive 2014/34/EU. Supplementary Certificates to such EC-Type Examination Certificates, and new issues of such certificates, may continue to bear the original certificate number issued prior to 20.04.2016.

(10) Compliance with the Essential Health and Safety Requirements has been assured by compliance with:

EN IEC 60079-0:2018

EN 60079-1:2014

EN 60079-11:2012

EN 50104:2019

EN 50271:2018

EN 50303:2000

EN 60079-29-1:2016+A1:2022+A11:2022

If the sign "X" is placed after the certificate number, it indicates that the product is subject to Specific Conditions of Use specified in the schedule to this certificate.

(11) The marking of the product shall include the following:



I M1

Ex ia I Ma

II 1G

Ex da ia IIC T3, T4 Ga

(12) This certificate is valid till: **30.09.2031**

Dipl. Ing. Lukáš Martinák

Representative of Notified Body No. 1026



Date of issue: 22.09.2026

Page: 1/3

This certificate is not valid without the signature and stamp of Notified Body No. 1026.
This certificate may only be reproduced in its entirety and without any change.



**Physical-Technical Testing Institute
Ostrava - Radvanice**

(13)

Schedule

(14) **Supplementary EU - Type Examination Certificate No. 23
to FTZÚ 08 ATEX 0340X**

(15) Description of the variation to the Product:

The subject of this supplementary certificate is:

- Modification of certified apparatus.
- Evaluation of changes according to the newest standards.
- Extension of certificate validity.

There are constructional changes – added alternative Bluetooth board. This modification has no negative effect on the type and level of the protection.

The product changes have been evaluated according to the standards, which are listed in the clause (10) of this supplement. The validity of certificate has been extended for next five years. The documentation has been updated and is listed in clause (19).

Applied Ex components:

Catalytic sensor MSA XCell Ex, Ex da ia IIC Ga, Ex ia Ma - certified by FTZÚ 09 ATEX 0123U

El. chem. sensor MSA XCell eChem, Ex ia IIC Ga, Ex ia Ma - certified by FTZÚ 09 ATEX 0223U

Parameters:

Charging contacts entity parameters: $U_m \leq 6.7 \text{ V}$ (for version with rechargeable battery)

Ambient temperature range: $-40 \text{ }^{\circ}\text{C}$ to $+50 \text{ }^{\circ}\text{C}$ for explosion safety

Degree of protection provided by enclosure: IP65

Temperature class depends of used battery:

T3 - with Varta High Energy 4906 or Energizer E91

T4 - with Panasonic CGR18650DA or Panasonic/Sanyo UR18650A or Duracell LR6, MN1500

Performance approvals for temperature range: -20°C to $+50^{\circ}\text{C}$ (recapitulation)

Gases CH_4 , C_3H_8 , H_2 , C_4H_{10} , C_5H_{12} in range 0-100% LFL for Group II by catalytic sensor.

Gas CH_4 in range 0-100% LFL for Group I by catalytic sensor.

Gas O_2 in range 0-25% (v/v) for inertization purposes only for Group I and II by electrochemical sensor.

Firmware version: v2.20.1

(16) Report Number: 08/0340/23

Dipl. Ing. Lukáš Martinák
Representative of Notified Body No. 1026



Date of issue: 22.09.2026

Page: 2/3

This certificate is not valid without the signature and stamp of Notified Body No. 1026.
This certificate may only be reproduced in its entirety and without any change.

Physical-Technical Testing Institute, s.p., Pikartská 1337/7, 716 00 Ostrava - Radvanice, Czech Republic
tel.: +420 595 223 111, +420 604 203 525, e-mail: ftzu@ftzu.cz, www.ftzu.cz



**Physical-Technical Testing Institute
Ostrava - Radvanice**

(13)

Schedule

(14) **Supplementary EU - Type Examination Certificate No. 23
to FTZÚ 08 ATEX 0340X**

(17) Specific Conditions of Use: (recapitulation)

1. The equipment shall only be charged by manufacturer's chargers in an ambient temperature from 0°C to +45°C and opened in a non-hazardous area.
2. The warm-up time for oxygen is up to 180 seconds.
3. In event of combustible sensor overrange the apparatus shall be exposed to fresh air for 20 minutes minimum and then the fresh air setup procedure shall be done.
4. The alarm set points are not applied for measuring oxygen inertization and it shall be taken into account.
5. The equipment enclosure includes accessible metal parts. The end user shall determine suitability in the specific application. The measured capacitance of the equipment with AA cells on the item: Battery fastening screw is 6 pF. The measured capacitance of the equipment with Li-ion cell on the item: D-ring is 26 pF and on the item: Pin for charging is 16 pF.
6. The antenna used for activation of the internal RFID tag with the RF radiation power shall not exceed 6 W for Group I and 2 W for Group IIC.
7. The EMC evaluation according to standard EN 50270:2015 is on manufacturer's responsibility.

(18) Essential Health and Safety Requirements:

Compliance with the Essential Health and Safety Requirements is covered by standards mentioned in clause (10) of this supplementary certificate.

(19) Drawings and Documents:

Number:	Issue:	Sheets:	Date:	Description:
10223097	4	9	09.2026	ADDENDUM Standards Compliance Certifications
SK3098-1184	12	98	14.01.2026	FTZÚ Approvals
SK3098-1185	2	1	01.07.2022	Label, ATEX Approvals, Altair 5X
10114801	10	68	09.2026	Operating Manual

Dipl. Ing. Lukáš Martinák
Representative of Notified Body No. 1026



Date of issue: 22.09.2026

Page: 3/3

This certificate is not valid without the signature and stamp of Notified Body No. 1026.
This certificate may only be reproduced in its entirety and without any change.

Physical-Technical Testing Institute, s.p., Pikartská 1337/7, 716 00 Ostrava - Radvanice, Czech Republic
tel.: +420 595 223 111, +420 604 203 525, e-mail: ftzu@ftzu.cz, www.ftzu.cz